

VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

APR 02 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		Datamaster cdm S/N <u>130125</u>	
Name <u>GARY J. TELLER</u>		ID# <u>2281</u>	Date <u>3-31-10</u>
A	Agency <u>AST</u>	Phone # <u>451.5100</u>	
Instrument Location <u>1979 Peyer Rd.</u>		Fbx <u>99701</u>	
Alco S/N <u>X172933</u>		Target Value <u>-081</u>	High Pressure <u>1025</u>
B	Alco Test Value Average		
	<u>-076</u>	<u>-077</u>	
	1 st Alco	2 nd Alco	
Signature <u>Gary J. Teller</u>			
(OVER)			


(Do Not write in the area below)

- I, Nita J. Bolz, after being first duly sworn, depose and state as follows:
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of May, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of Datamaster cdm

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Datamaster cdm S/N

130125

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

MARCH 31, 2010

OPERATOR'S NAME:

TELLEP GARY J

OPERATOR'S NUMBER: 2281

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

N

O.L. #: N

DEPT/AGENCY: FBK0

CASE/REPORT: N

TEST TYPE: U

ALCO TARGET VALUE: .081

ALCO S/N: X172933

--- BREATH ANALYSIS ---

.081 ADJUSTED FOR 29.28 in

ALCO TARGET .079 05:44

BLANK TEST .000 05:45

INTERNAL STANDARD VERIFIED 05:45

ALCO TU 29.29 in .076 05:45

BLANK TEST .000 05:46

SUBJECT SAMPLE .000 05:47

BLANK TEST .000 05:47

ALCO TU 29.28 in .077 05:48

BLANK TEST .000 05:49

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130125

MARCH 31, 2010

TIME 05:50

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 43c

BREATH TUBE: 42c

BAROMETER: 29.28 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKL MNOPQRSTUVWXYZ [\] ^ _ ` abcdefghijklmno
pqrstuvwxyz{|}~